



hs-CRP+CRP Fast Test Kit (Immunofluorescence Assay)

IF2003 for Getein1600
IF4003 for Getein1200
IF5003 for Getein1160
IF3003 for Getein1180
IF1003 for Getein1100
IF6003 for Getein208

REF

User Manual

INTENDED USE

hs-CRP+CRP Fast Test Kit (Immunofluorescence Assay) is intended for *in vitro* quantitative determination of C-reactive protein (CRP) in human serum, plasma, whole blood or fingertip blood samples. Measurement of CRP is useful for the detection and evaluation of infection, tissue injury and inflammatory disorders. Measurement of high sensitivity CRP (hs-CRP), when used in conjunction with traditional clinical laboratory evaluation of acute coronary syndromes (ACS), may be useful as an independent marker of prognosis for recurrent events in patients with stable coronary disease or ACS.

SUMMARY

C-reactive protein is an acute-phase reactant that precipitated with Pneumococcal C-polysaccharide, and is a non-specific immune response component. CRP has wide distribution in our body, and is an acute phase protein produced in the liver in response to microbial infection or tissue injury, it measures general levels of inflammation in the body, and the hs-CRP can be used to detect lower concentrations of CRP in serum or plasma. Studies revealed hs-CRP levels seem to be correlated with Atherosclerosis and Acute Myocardial Infarction. And the hs-CRP is an inflammation "marker" for ACS patient and is helpful for primary prevention and risk assessment of cardiovascular disease. Its combination with the ratio of total cholesterol to HDL-C is more accurate than other risk factors in predicting cardiovascular disease.

The American Heart Association and US Centers for Disease Control and Prevention have advocated hs-CRP as a predictor of cardiovascular disease (CVD) to define risk groups: less than 1.0 mg/L indicates low risk, 1.0 to 3.0 mg/L means moderate risk, and the amount above 3.0 mg/L (lower than 10 mg/L) strongly suggests a high risk of CVD. Moreover, higher CRP levels are found in late pregnant women, mild inflammation and viral infections (10–40 mg/L), active inflammation, bacterial infection (40–200 mg/L), severe bacterial infections and burns (>200 mg/L).

PRINCIPLE

The test uses an anti-human CRP monoclonal antibody

conjugated with fluorescence latex and another anti-human CRP monoclonal antibody coated on the test line. After the sample has been applied to the test strip, the fluorescence latex-labelled anti-human CRP monoclonal antibody binds with the CRP in sample and forms a marked antigen-antibody complex. This complex moves to the test card detection zone by capillary action. Then marked antigen-antibody complex is captured on the test line by the anti-human CRP monoclonal antibody. The fluorescence intensity of the test line increases in proportion to the amount of CRP in sample.

Then insert test card into Getein1100/Getein1160/Getein1180 Immunofluorescence Quantitative Analyzer/Getein208 Hand-held Integrated System/automatically inserted by Getein1200/Getein1600 Immunofluorescence Quantitative Analyzer (hereinafter referred to as Getein1100, Getein1160, Getein1180, Getein208, Getein1200 and Getein1600), the concentration of CRP in sample will be measured and displayed on the screen. The value will be stored in Getein1100/Getein1160/Getein1180/Getein208/Getein1200/Getein1600 and available for downloading. The result can be easily transmitted to the laboratory or hospital information system.

CONTENTS

1. A kit for Getein1100/Getein1160/Getein1180/Getein208 contains:

- Package specifications: 25 tests/box, 10 tests/box
- 1) hs-CRP+CRP test card in a sealed pouch with desiccant
 - 2) Capillary pipet
 - 3) Sample diluent
 - 4) User manual: 1 piece/box
 - 5) SD card: 1 piece/box

2. A kit for Getein1200/Getein1600 contains:

- Package specifications: 2x24 tests/kit, 2x48 tests/kit
- 1) Sealed cartridge with 24/48 Getein hs-CRP+CRP test cards
 - 2) User manual: 1 piece/box
- Materials required for Getein1200/Getein1600:

- 1) Sample diluent: 1 bottle/box
- 2) Box with pipette tips: 96 tips/box
- 3) Mixing plate: 1 piece/box

3. Sample diluent composition:

Phosphate buffered saline, proteins, detergent, preservative, stabilizer.

4. A test card consists of:

A plastic shell and a reagent strip which is composed of a sample pad, nitrocellulose membrane (one end of the membrane is coated with a fluorescence latex-labelled anti-human CRP monoclonal antibody), the test line is coated with another anti-human CRP monoclonal antibody and the control line is coated with rabbit anti-mouse IgG antibody), absorbent paper and liner.

Note: Do not mix or interchange different batches of kits.

APPLICABLE CASE

Getein1100 Immunofluorescence Quantitative Analyzer
Getein1180 Immunofluorescence Quantitative Analyzer
Getein208 Hand-held Integrated System
Getein1600 Immunofluorescence Quantitative Analyzer
Getein1160 Immunofluorescence Quantitative Analyzer
Getein1200 Immunofluorescence Quantitative Analyzer

STORAGE AND STABILITY

Store the test card at 4–30°C with a valid period of 24 months. Use the test card for Getein1100/Getein1160/Getein1180/Getein208 within 1 hour once the foil pouch is opened. For test card of Getein1200/Getein1600: if the cartridge is opened, it could be stable within 24 hours once exposed to air. If the test cards can't be used up at a time, please put the cartridge back to the foil pouch and reseal along the entire edge of zip-seal. The remaining test cards should be used up within 7 days.

PRECAUTIONS

1. For *in vitro* diagnostic use only.
2. Do not use the kit beyond the expiration date.
3. Do not use the test card if the foil pouch or the cartridge is damaged.
4. Do not open pouches or the cartridge until ready to perform the test.
5. Do not reuse the test card.
6. Do not reuse the pipet.
7. Handle all specimens as potentially infectious. Proper handling and disposal methods should be followed in accordance with local regulations.
8. Carefully read and follow user manual to ensure proper test performance.

SPECIMEN COLLECTION AND PREPARATION

1. This test can be used for **serum, plasma, whole blood and fingertip blood samples. Heparin, sodium citrate and EDTA** can be used as the anticoagulant for plasma, whole blood and fingertip blood. Samples should be free of hemolysis.
2. Suggest using serum or plasma for better results.
3. If testing is delayed, serum and plasma samples may be stored up to 7 days at 2–8°C or stored at -20°C for 6 months before testing (whole blood sample may be stored up to 3 days at 2–8°C).
4. Refrigerated or frozen sample should reach room temperature and be homogeneous before testing. Avoid multiple freeze-thaw cycles.
5. Do not use heat-inactivated samples.
6. **SAMPLE VOLUME (for Getein1100/Getein1160/Getein1180/Getein208): 10 μ L.**

TEST PROCEDURE

1. Collect specimens according to user manual.
2. Test card, sample and reagent should be brought to room

temperature before testing.

For Getein1100:

3. Confirm SD card lot No. in accordance with test kit lot No.. Perform "SD card" calibration when necessary.
4. Remove the test card from the sealed pouch immediately before use. Label the test card with patient or control identification.
5. Put the test card on a clean table, horizontally placed.
6. Using sample transfer pipette, deliver **10 μ L** of sample into one tube of sample diluent, mix gently and thoroughly. Then drop **100 μ L** of sample mixture into the sample port on the test card (for disposable capillary pipet using, please refer to the directions in the package).
7. **Reaction time: 3 minutes.** Insert the test card into Getein1100 and press "ENT" button or click on "Start" icon (for Android Getein1100) after reaction time is elapsed. The result will be shown on the screen and printed automatically.

For Getein1160/Getein1180:

8. Confirm SD card lot No. in accordance with test kit lot No.. Perform "SD card" calibration when necessary.
9. Enter testing interface of Getein1160/Getein1180.
10. Remove the test card from the sealed pouch immediately before use. Label the test card with patient or control identification.
11. Put the test card on a clean table, horizontally placed.
12. Using sample transfer pipette, deliver **10 μ L** of sample into one tube of sample diluent, mix gently and thoroughly. Then drop **100 μ L** of sample mixture into the sample port on the test card.
13. **Reaction time: 3 minutes.** Insert the test card into Getein1160/Getein1180 immediately after sample loading. The analyzer will count down the reaction time and automatically test the card after reaction time is elapsed. The result will be shown on the screen and printed automatically.

For Getein208:

14. Long press the Power Button to start the analyzer
15. The system will enter (Test) menu.
16. Insert the MEMO memory chip which is with the same batch number as the test card.
17. Select (Test) menu, press (OK) to enter [Read Calibration Card] interface.
18. Press (OK) to automatically obtain the test item, batch number, serial number and sampling volume. Select the sample type by pressing <or> buttons.
19. Press (OK). The screen then prompts [Insert test card] and starts counting down from 60 sec. Insert test card within the 60 sec.

Note: Do not move the test card after it is inserted.

20. Add sample within 120 sec when the screen prompts [Wait for sample]. Then draw **10 μ L** of sample and drop it into 1000 μ L of sample diluent. Then drop **60 μ L** of sample mixture into the sample port on the test card.
21. After sample adding, the system starts react-time countdown automatically.

22. After the countdown is over, the system starts testing automatically. Please check and record test results then.

Note: Test results are saved automatically in the system.

23. Long Press (OK) to return to the main interface. Take out and discard the test card.

For Getein1200/Getein1600:

24. Each cartridge for Getein1200/Getein1600 contains a specific RFID card which can calibrate automatically.

25. Place the sample diluent at the correct position in Getein1200/Getein1600.

26. Place samples in the designed area of the sample holder, insert the holder and select the right test item, Getein1200/Getein1600 will do the testing and print the result automatically.

Notes:

1. It is required to perform "SD card" calibration when using a new batch of kits for Getein1100/Getein1160/Getein1180/Getein208.
2. It is suggested to calibrate once for one batch of kits for Getein1100/Getein1160/Getein1180/Getein208.
3. Make sure the test card and the sample insertion is correct and complete.

TEST RESULTS

Getein1100/Getein1160/Getein1180/Getein208/Getein1200/Getein1600 can scan the test card automatically and display the result on the screen. For additional information, please refer to the user manual of Getein1100/Getein1160/Getein1180/Getein208/Getein1200/Getein1600.

EXPECTED VALUE

hs-CRP: The expected normal value for hs-CRP was determined by testing samples from 500 apparently healthy individuals. The 95th percentile of the concentration for hs-CRP is 3.0 mg/L. (The probability that hs-CRP value of a normal person below 3.0 mg/L is 95%.)

CRP: The expected normal value for CRP was determined by testing samples from 500 apparently healthy individuals. The 95th percentile of the concentration for CRP is 10.0 mg/L. (The probability that CRP value of a normal person below 10.0 mg/L is 95%.)

It is recommended that each laboratory establish its own expected values for the population it serves.

PERFORMANCE CHARACTERISTICS

Measuring Range	0.5–200.0 mg/L
Lower Detection Limit	≤0.5 mg/L
Within-Run Precision	≤10%
Between-Run Precision	≤15%

LIMITATIONS

1. As with all diagnostic tests, a definitive clinical diagnosis should not be made based on the result of a single test. The

test results should be interpreted considering all other test results and clinical information such as clinical signs and symptoms.

2. Samples containing interferents may influence the results. The table below lists the maximum allowance of these potential interferents.














Interferent	Hemoglobin	Triglyceride	Bilirubin
Concentration (Max)	10 g/L	10 g/L	0.2 g/L

REFERENCES

1. Danesh J, Whincup P, Wslker M, et al. Low grade inflammation and coronary heart disease: prospective study and updated meta-analysis. *BJM* 2000; 321:199–204.
2. Rifai N, Ridker PM. Proposed cardiovascular risk assessment algorithm using high-sensitivity C-reactive protein and lipid screening. *Clin Chem* 2001; 47:28–30.
3. EN ISO 18113-1:2011 In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 1: Terms, definitions and general requirements.
4. EN ISO 18113-2:2011 In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 2: In vitro diagnostic reagents for professional use.

DESCRIPTION OF SYMBOLS USED

The following graphical symbols used in or found on hs-CRP+CRP Fast Test Kit (Immunofluorescence Assay) are the most common ones appearing on medical devices and their packaging. They are explained in more detail in the European Standard EN ISO 15223-1:2021.

Key to symbols used			
	Manufacturer		Use-by date
	Do not re-use		Date of manufacture
	Consult instructions for use or consult electronic instructions for use		Batch code
	Temperature limit		In vitro diagnostic medical device
	Contains sufficient for <n> tests		Authorized representative in the European Community/European Union
	CE mark		Do not use if package is damaged and consult instructions for use
	Catalogue number		

Thank you for purchasing hs-CRP+CRP Fast Test Kit (Immunofluorescence Assay). Please read this user manual carefully before operating to ensure proper use.

Version: WIF07-S-11



Getein Biotech, Inc.

Add.: No.9 Bofu Road, Luhe District, Nanjing, 211505, China

Tel: +86-25-68568508

Fax: +86-25-68568500

E-mail: tech@getein.com.cn

overseas@getein.com.cn

Website: www.getein.com



Lotus NL B.V.

Add.: Koninng Julianaplein 10, 1e Verd, 2595AA, The Hague, Netherlands.

E-mail: peter@lotusnl.com

Tel: +31644168999