



# HBsAg Fast Test Kit (Immunofluorescence Assay)

## User Manual



IF1064 for Getein1100  
IF3064 for Getein1180  
IF2064 for Getein1600

## INTENDED USE

HBsAg Fast Test Kit (Immunofluorescence Assay) is intended for *in vitro* quantitative determination of hepatitis B surface antigen (HBsAg) in serum or plasma samples. This test is suitable for blood supply screening and clinical diagnosis of hepatitis B virus infection.

## SUMMARY

Hepatitis B is an infection of the liver caused by the Hepatitis B Virus (HBV). HBV is transmitted by exposure to infectious blood or body fluids (e.g. saliva, semen). Forms of transmission include unprotected sexual activity, blood transfusion, mother-to-infant transmission, or consuming contaminated food.

The average incubation period for HBV infection is 6 to 8 weeks (ranges from 1 to 6 months). Common clinical symptoms include malaise, fever, gastroenteritis, and icterus. In adults, 90% to 95% of patients with HBV infection completely recover from acute illness and clear the virus, approximately 5% to 10% of patients with HBV become chronic carriers. HBsAg is a coat protein of hepatitis B virus. It is composed of mixed polypeptides, containing lipids, sugars and proteins. It has strong resistance to low temperature and is not contagious itself. However, its appearance is often accompanied by the presence of hepatitis B virus, so it is a sign that it has been infected with hepatitis B virus. It can be present in the patient's blood, saliva, milk, sweat, tears, nasopharyngeal secretions, semen and vaginal secretions. It is estimated that over 300 million people worldwide are chronic carriers of the virus.

HBsAg usually appears 1 to 2 weeks after infection with hepatitis B virus. Most patients with acute hepatitis B can turn negative in the early stage of the disease, and this indicator can be positive for patients with chronic hepatitis B. This kit is for the detection of HBsAg and is suitable for blood supply screening and auxiliary diagnosis of clinical hepatitis B virus infection. HBsAg is observed in persons with acute and chronic hepatitis B infections.

## PRINCIPLE

This test uses an anti-human HBsAg monoclonal antibody conjugated

with fluorescence latex and another anti-human HBsAg monoclonal antibody coated on the test line. After the sample has been applied to the test strip, the fluorescence latex-labeled anti-human HBsAg monoclonal antibody binds with the HBsAg in sample and forms marked antigen-antibody complex. The complex moves to the detection zone by capillary action. Then marked antigen-antibody complex is captured on the test line by the anti-human HBsAg monoclonal antibody. The fluorescence intensity of the test line increases in proportion to the amount of HBsAg in the sample.

Then insert test card into Getein1100/Getein1180 Immunofluorescence Quantitative Analyzer/Automatically inserted by Getein1600 Immunofluorescence Quantitative Analyzer (hereafter referred to as Getein1100, Getein1180 and Getein1600), the concentration of HBsAg in the sample will be measured and displayed on the screen. The value will be stored in Getein1100/Getein1180/Getein1600 and available for downloading. The result can be easily transmitted to the laboratory or hospital information system.

## CONTENTS

### 1. A kit for Getein1100/Getein1180 contains:

Package specifications: 25 tests/box, 10 tests/box

- 1) Getein HBsAg test card in a sealed pouch with desiccant
- 2) Disposable pipet
- 3) User manual: 1 piece/box
- 4) SD card: 1 piece/box

### 2. A kit for Getein1600 contains:

Package specifications: 2x24 tests/kit, 2x48 tests/kit

Sealed cartridge with 24/48 Getein HBsAg test cards

User manual: 1 piece/box Materials required for Getein1600:

- 1) Sample diluent: 1 bottle/box
- 2) Box with pipette tips: 96 tips/box
- 3) Mixing plate: 1 piece/box

### 3. Sample diluent composition:

Phosphate buffered saline, proteins, detergent, preservative, stabilizer.

### 4. A test card consists of:

A plastic shell and a reagent strip which is composed of a sample pad, nitrocellulose membrane (one end of the membrane is coated with fluorescence latex-labeled anti-human HBsAg monoclonal antibody I, the test line is coated with another anti-human HBsAg monoclonal antibody II, and the control line is coated with rabbit anti-mouse IgG antibody), absorbent paper and liner.

**Note:** Do not mix or interchange different batches of kits.

## APPLICABLE DEVICE

Getein1100 Immunofluorescence Quantitative Analyzer  
Getein1180 Immunofluorescence Quantitative Analyzer  
Getein1600 Immunofluorescence Quantitative Analyzer

## STORAGE AND STABILITY

Store the test card at 4~30°C with a valid period of 24 months. Use the test card for Getein1100/Getein1180 within 1 hour once the foil pouch is opened.

For test card of Getein1600: If the cartridge is opened, it could be stable within 24 hours once exposure to air. Please put the cartridge back to the foil pouch and reseal along the entire edge of zip-seal. The remaining test cards should be used up within 7 days. Store the sample diluent at 0~30°C with a valid period of 24 months.

## PRECAUTIONS

1. For *in vitro* diagnostic use only.
2. Do not use the kit beyond the expiration date.
3. Do not use the test card if the foil pouch is damaged.
4. Do not open pouches until ready to perform the test.
5. Do not reuse the test card.
6. Do not reuse the pipet.
7. Handle all specimens as potentially infectious. Proper handling and disposal methods should follow by local regulations.
8. Carefully read and follow the user manual to ensure an appropriate test performance.

## SPECIMEN COLLECTION AND PREPARATION

1. This test can be used for **serum and plasma samples**. **Heparin, sodium citrate and EDTA** can be used as the anticoagulant for plasma. Samples should be free of hemolysis.
2. The test should be performed within 4 hours after blood collection.
3. If testing will be delayed, serum and plasma samples may be stored up to 5 days at 2~8°C or stored at -20°C for 6 months before testing.
4. Refrigerated or frozen sample should be reached to room temperature and be homogeneous before testing. Avoid multiple freeze-thaw cycles.
5. Do not use heat-inactivated or hemolysis samples.
6. **SAMPLE VOLUME**(for Getein1100/Getein1180) : **100  $\mu$ L**

## TEST PROCEDURE

1. Collect specimens according to user manual.
2. Test card, sample and reagent should reach to room temperature before testing.
3. For Getein1100.  
Confirm SD card lot No. in accordance with test kit lot No. Perform "SD card" calibration when necessary.
4. Remove the test card from the sealed pouch immediately before use. Label the test card with patient or control identification.
5. Put the test card on a clean table, horizontally placed.
6. Using sample transfer pipette, deliver **100  $\mu$ L** of sample into the sample port on the test card.

7. **Reaction time: 15 minutes.** Insert the test card into Getein1100 and click on "Start" icon after reaction time is elapsed. The result will be shown on the screen and printed automatically.

**For Getein1180:**

8. Confirm SD card lot No. in accordance with test kit lot No. Perform "SD card" calibration when necessary.
9. Enter testing interface of Getein1180.
10. Remove the test card from the sealed pouch immediately before use. Label the test card with patient or control identification.
11. Put the test card on a clean table, horizontally placed.
12. Using sample transfer pipette, deliver **100 µL** of sample into the sample port on the test card.
13. **Reaction time: 15minutes.** Insert the test card into Getein 1180 immediately after sample loading. The analyzer will count down the reaction time and automatically test the card after reaction time is elapsed. The result will be shown on the screen and printed automatically.

**For Getein1600:**

14. Each cartridge for Getein1600 contains a specific RFID card which can calibrate automatically.
15. Put the sample diluent at the correct position in Getein1600.
16. Place samples in the designated area of the sample holder, insert the holder and select the right test item, Getein1600 will do the testing and print the result automatically.

**Notes:**

1. It is required to perform "SD card" calibration when using a new batch of kits.
2. It is suggested to calibrate once for one batch of kits for Getein-1100/Getein1180.
3. Make sure the test card and the sample insertion is correct and complete.

**TEST RESULTS**

1. Getein1100/Getein1180/Getein1600 can scan the test card automatically and display the result on the screen. For additional information, please refer to the user manual of Getein1100/Getein1180/Getein1600.
2. Samples with concentration <1.00 IU/mL are considered negative and no further testing is required.
3. Samples with concentration ≥1.00 IU/mL are considered positive. All positive samples that are initially tested should be retested twice. If both retests are negative, the sample must be considered HBsAg negative. If any of the retest values is positive, it is considered HBsAg positive.
4. Due to different methodologies or antibody specificity, there may be deviations between the test results of different manufacturers, so they can't be compared directly.
5. Dilute the sample which concentration is higher than the upper limit with negative samples, and the dilution ratio should be less than 80 times.

**EXPECTED VALUE**

The expected normal value for HBsAg was determined by testing samples from 460 apparently healthy individuals. The reference range of HBsAg is 1.00 IU/mL calculated by using normal distribution methods giving a level of confidence of approximately 99%. It is recommended that each laboratory establish its own expected values for the population it serves.

**PERFORMANCE CHARACTERISTICS**

Measuring Range	1.00~100.00 IU/mL
Lower Detection Limit	≤ 1.00 IU/mL
Within-run Precision	≤ 10%
Between-run Precision	≤ 15%

**LIMITATIONS**

As with all diagnostic tests, a definitive clinical diagnosis should not be made based on the result of a single test. The test results should be interpreted considering all other test results and clinical information such as clinical signs and symptoms.

**REFERENCES**

1. Godoy BA, Gomes-Gouvea MS, Zaganel-Oliveira M, et al. High prevalence of HBV/A1 subgenotype in native south Americans may be explained by recent economic developments in the Amazon. *Infection Genetics and Evolution* 2016, 43: 354-363.
2. Nguyen LH, Hoang J, Nguyen NH, et al. Ethnic differences in incidence of hepatitis B surface antigen seroclearance in a real-life multicenter clinical cohort of 4737 patients with chronic hepatitis B infection. *Alimentary Pharmacology & Therapeutics* 2016, 44(4): 390-399.
3. Verma AK, Sharma S, Gupta P, et al. Vitamin B-12 Grafted Layer-by-Layer Liposomes Bearing HBsAg Facilitate Oral Immunization: Effect of Modulated Biomechanical Properties. *Molecular Pharmaceutics* 2016, 13(7): 2531-2542.
4. Ork V, Woodring J, Shafiqul MD, Annemarie H, et al. Hepatitis B surface antigen seroprevalence among pre- and post-vaccine cohorts in Cambodia, 2017 *Vaccine*, 2019, 37(35).
5. Mathias EE, Mbouamba YB, Esemu S, et al. The prevalence of HBsAg, knowledge and practice of hepatitis B prevention among pregnant women in the Limbe and Muyuka Health Districts of the South West region of Cameroon: a three-year retrospective study. *The Pan African medical journal*, 2019, 32.
6. Cai QY, Liu HY, Han WH, et al. Maternal HBsAg carriers and adverse pregnancy outcomes: A hospital-based prospective cohort analysis. *Journal of viral hepatitis*, 2019, 26(8).
7. Chen L, Shi JJ, Lu ZH, et al. Baseline HBsAg levels associated with HBsAg loss in HBeAg-negative chronic hepatitis B infection with persistently normal alanine aminotransferase. *Clinics and Research in Hepatology and Gastroenterology*, 2019, 43(3).

**DESCRIPTION OF SYMBOLS USED**

The following graphical symbols used in or found on HBsAg Fast Test Kit (Immunofluorescence Assay) are the most common ones appearing on medical devices and their packaging. They are explained in more details in the European Standard EN ISO 15223-1:2016.

Key to symbols used			
	Manufacturer		Use-by date
	Do not re-use		Date of manufacture
	Consult instructions for use		Batch code
	Temperature limit		In vitro diagnostic medical device
	Contains sufficient for <n> tests		Do not use if package is damaged
	Catalogue number		

Thank you for purchasing HBsAg Fast Test Kit (Immunofluorescence Assay). Please read this user manual carefully before operating to ensure proper use.

Version: WIF64-S-04



Getein Biotech, Inc.  
 Add: No.9 Bofu Road, Luhe District, Nanjing, 211505, China  
 Tel: +86-25-68568508  
 Fax: +86-25-68568500  
 E-mail: tech@getein.com.cn  
 overseas@getein.com.cn  
 Website: www.getein.com